

# SIEMENS MAGNETOM Verio syngo MR B17

\\USER\Dr. Tarr\John\5000Scenes\SpinEchoFieldMap\_PA\_CSIX\_Sess-X

TA: 0:37 PAT: Off Voxel size: 2.0x2.0x2.0 mm Rel. SNR: 1.00 USER: cmrr\_mbep2d\_se

## Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	Off
Start measurements	single

## Routine

Slice group 1	
Slices	69
Dist. factor	0 %
Position	R4.0 P18.0 H22.0
Orientation	T > C-5.0
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	9240 ms
TE	81.6 ms
Multi-band accel. factor	1
Filter	Prescan Normalize
Coil elements	HEA;HEP

## Contrast

MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Disabled
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	3
Delay in TR	0 ms
Multiple series	Off

## Resolution

Base resolution	106
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Auto (CP)
Distortion Corr.	Off
Unfiltered images	On
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

## Geometry

Multi-slice mode	Interleaved
Series	Interleaved

## Special sat.

None

## System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Auto Coil Select	Default

Shim mode	Advanced
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R4.0 P18.0 H22.0
Orientation	T > C-5.0
Rotation	0.00 deg
R >> L	212 mm
A >> P	212 mm
F >> H	138 mm

## Physio

1st Signal/Mode	None
-----------------	------

## BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	15
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Motion correction	Off
Spatial filter	Off

## Sequence

Introduction	Off
Contrasts	1
Bandwidth	1814 Hz/Px
Free echo spacing	Off

## SIEMENS MAGNETOM Verio syngo MR B17

Echo spacing	0.72 ms
EPI factor	106
RF pulse type	Normal
Gradient mode	Fast
SENSE1 coil combine	Off
Invert RO/PE polarity	On
Force equal slice timing	Off
FFT scale factor	1.00
Physio recording	Off
Triggering scheme	Standard