

SIEMENS MAGNETOM Verio syngo MR B17

\\USER\Dr. Tarr\John\5000Scenes\DSI_MB3-253d_PA

TA: 0:40 PAT: Off Voxel size: 2.0x2.0x2.0 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_diff

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	69
Dist. factor	0 %
Position	R4.0 P17.0 H27.0
Orientation	Transversal
Phase enc. dir.	A >> P
Rotation	0 deg
Phase oversampling	0 %
FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2 mm
TR	3981 ms
TE	121 ms
Multi-band accel. factor	3
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	90 deg
Refocus flip angle	180 deg
Fat suppr.	Fat sat.
Grad. rev. fat suppr.	Enabled
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	106
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Auto (CP)
Distortion Corr.	Off
Prescan Normalize	On
Raw filter	Off
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved

Special sat.

None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Auto Coil Select	Default
Shim mode	Advanced
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R4.0 P17.0 H27.0
Orientation	Transversal
Rotation	0.00 deg
R >> L	212 mm
A >> P	212 mm
F >> H	138 mm

Physio

1st Signal/Mode	None
-----------------	------

Diff

Diffusion mode	Free
Diff. weightings	1
b-value	4000 s/mm²
Diff. weighted images	On
Trace weighted images	Off
Average ADC maps	Off
Individual ADC maps	Off
FA maps	Off
Mosaic	On
Tensor	Off
Noise level	40
Diff. directions	253

Sequence

Introduction	Off
Bandwidth	1684 Hz/Px
Free echo spacing	Off
Echo spacing	0 ms
EPI factor	106
Gradient mode	Fast
RF spoiling	Off
Excite pulse duration	3840 us
Refocus pulse duration	6400 us
Diffusion Scheme	Bipolar
Single-band images	On
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
Time-shifted MB RF	Off
SENSE1 coil combine	On
Invert RO/PE polarity	On

SIEMENS MAGNETOM Verio syngo MR B17

PF omits higher k-space	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1
Physio recording	Off