

SIEMENS MAGNETOM Verio syngo MR B17

\\USER\Dr. Tarr\John\5000Scenes\BOLD_CSIX_Sess-X_Run-X

TA: 6:38 PAT: Off Voxel size: 2.0x2.0x2.0 mm Rel. SNR: 1.00 USER: cmrr_mbep2d_bold

Properties

Prio Recon	Off
Before measurement	
After measurement	
Load to viewer	On
Inline movie	Off
Auto store images	On
Load to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Start measurement without further preparation	On
Wait for user to start	On
Start measurements	single

Routine

Slice group 1	
Slices	69
Dist. factor	0 %
Position	R4.0 P18.0 H22.0
Orientation	T > C-5.0
Phase enc. dir.	A >> P
Rotation	0.00 deg
Phase oversampling	0 %
FoV read	212 mm
FoV phase	100.0 %
Slice thickness	2.00 mm
TR	2000 ms
TE	30.0 ms
Multi-band accel. factor	3
Filter	Prescan Normalize
Coil elements	HEA;HEP

Contrast

MTC	Off
Magn. preparation	None
Flip angle	79 deg
Fat suppr.	Fat sat.
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	194
Delay in TR	0 ms
Multiple series	Off

Resolution

Base resolution	106
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off
PAT mode	None
Matrix Coil Mode	Auto (CP)
Distortion Corr.	Off
Unfiltered images	On
Prescan Normalize	On
Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry

Multi-slice mode	Interleaved
Series	Interleaved
Special sat.	None

System

Body	Off
HEP	On
HEA	On
Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Auto Coil Select	Default
Shim mode	Advanced
Adjust with body coil	Off
Confirm freq. adjustment	Off
Assume Silicone	Off
? Ref. amplitude 1H	0.000 V
Adjustment Tolerance	Auto
Adjust volume	
Position	R4.0 P18.0 H22.0
Orientation	T > C-5.0
Rotation	0.00 deg
R >> L	212 mm
A >> P	212 mm
F >> H	138 mm

Physio

1st Signal/Mode	None
-----------------	------

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Starting ignore meas	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off

Sequence

Introduction	Off
--------------	-----

SIEMENS MAGNETOM Verio syngo MR B17

Contrasts	1
Bandwidth	1814 Hz/Px
Flow comp.	No
Free echo spacing	Off
Echo spacing	0.72 ms
<hr/>	
EPI factor	106
Gradient mode	Fast
RF spoiling	Off
<hr/>	
Excite pulse duration	8200 us
Single-band images	Off
MB LeakBlock kernel	Off
MB dual kernel	Off
MB RF phase scramble	Off
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Force equal slice timing	Off
Online multi-band recon.	Online
FFT scale factor	1.00
Physio recording	Legacy
Triggering scheme	Standard